



Food Safety

3M Center
St. Paul, MN 55144-1000
www.3M.com/foodsafety

Certificate of Analysis

Product Identification

3M Catalog Number	L25MLK
Product Description	3M™ Bovine Total Milk Protein Rapid Kit - 25 Tests per Kit, 1 Kit per Case
3M Stock Number	70-2011-7553-9
3M Kit Lot Number	21172
3M Buffer Lot Number	21172
Kit Expiration Date	7/6/2023

Performance Testing*

Negative Controls:

Sample	Expected Results	Pass/Fail
Matrix Testing	Negative, ≥ 24	Pass

Positive Controls:

Sample	Expected Results	Pass/Fail
Matrix Testing	Positive, ≥ 24	Pass


*A minimum of four matrices are used for performance testing. All results are interpreted by eye, holding the strip an arm's length away. If the control, hook and test lines are visible the test is considered positive. If only the control and hook lines are visible the test is considered negative. If either the control or hook lines aren't visible the test is considered invalid. Positive controls are spiked at 2ppm. This letter certifies that the product described above has been tested and found to comply with 3M's product performance specifications applicable to subject device at time of manufacture. Visit the 3M Food Safety website, www.3m.com/foodsafety for additional information.

Prepared By:

Signed:

Name:

Date:


Patrick McMonagle QA/QC Representative
7/9/2021