

Certificate of Quality

Product Description: Swab-Sampler w/ 10ml Buffered Peptone Water Broth (Formerly RediSwab)
Product Number: RS96010BPW
3M ID: 70200750878
ERP #: 7100010351
Lot Number: 33D835
Expiration Date: 2022-08-16

Release Specifications:

Representative samples of this lot were examined for clarity, pH, fill volume, and survival of low levels of organism inoculated into the product.

Clarity/precipitate: Clear / None

pH: (specification is 7.2 +/- 0.2): 7.2

Avg. Fill Volume: (specification is 10.0 +/- 0.5): 10.1 ml

Survival of Inoculated Organisms:

Survival of test organisms is evaluated by inoculating low levels of organisms into the product and holding @ 4° C for 1 hour.

Microorganism Panel	Log Colony Forming Units/ml	
	Time 0	One Hour
<i>Escherichia coli</i> ATCC 25922	2.10	2.07
<i>Escherichia coli</i> ATCC 35218	2.03	2.02
<i>S. typhimurium</i> ATCC 14028	2.14	2.13


The results met specification for survival of test organisms in this product.

Terminal Irradiation Processing Parameters

This product has been subjected to Gamma irradiation as a terminal process. The minimum and maximum delivered met established specifications for this product.

Approval Status:

This lot was found to be satisfactory for all quality characteristics measured and was released for commercial sale.

Signed:	
Name:	Lola K. Johnson
Date:	02/22/2021 10:54 AM

