



Health Care

3M Center

St. Paul, MN 55144-1000

Certificate of Analysis & Sterility

Product Identification

| | |
|---------------------|----------------------|
| 3M Catalog Number | 6486 |
| Product Description | 3M™ 1mL Pipette Tips |
| 3M Stock Number | 70-2006-9118-9 |
| 3M Lot Number | 501242680 |
| Expiration Date | 09/2026 |

This is to certify that the product described above has been manufactured in accordance with ISO 9001/ISO 13485 and has passed all relevant Quality Assurance procedures related to this product.

The sterilization process is electron beam (beta irradiation) in accordance with EN 552:1994 and EN 46002:1996.

This letter certifies that the product listed on this certificate has been manufactured according to 3M Health Care specifications.

Prepared By:

QA/QC Representative: Patrick McMonagle

Date: 10/27/2021

QA/QC Representative