



Petrifilm™

Created by Authorized Personnel: Jacob Billings, 2023-08-23

Manufacture Date: 2023-07-17

Expiration Date: 2025-01-14

Product Manufacturing Certificate

Certificate of Analysis

Product: 6407, 6417 or 6445 3M™ Petrifilm™ Yeast and Mold Count Plates
 Batch: 33R8W4
 Stock Number: 70-2005-6124-2 or 70-2005-9015-9 or 70-2007-7904-2
 ERP Number: 7100039379 or 7100039423 or 7100047854

Organism Tested	Minimum Growth	Result
<i>Aspergillus niger</i> , M6	Productivity Ratio \geq 0.7	Pass
<i>Paecilomyces</i> species, M10	Productivity Ratio \geq 0.7	Pass
<i>Botrytis</i> species, M97	Productivity Ratio \geq 0.7	Pass
<i>Hansenula anomala</i> , Y28	Productivity Ratio \geq 0.7	Pass
<i>Saccharomyces cerevisiae</i> , Y41	Productivity Ratio \geq 0.7	Pass
<i>Enterococcus faecalis</i>	No Growth	Pass
<i>Acinetobacter</i> species	No Growth	Pass

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through and independent agency.

3M Company
 3M Center, Building 275-5W-05
 St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
 Printed in USA © 3M 2022. All rights reserved.

Version 7

YM
 Yeast and Mold Count Plate