



Petrifilm™

Created by Authorized Personnel: Jacob Billings, 2023-05-16

Manufacture Date: 2023-04-30

Expiration Date: 2024-10-28

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ Lactic Acid Bacteria Count Plates 6461 or 6462
 Batch: 33PCP3
 Stock Number: 70-2007-9029-6 or 70-2007-9030-4
 ERP Number: 7100088605 or 7100088604

Organism Tested	Growth Specification	Result
<i>Leuconostoc mesenteroides</i> subsp. <i>mesenteroides</i>	Productivity Ratio ≥ 0.5	Pass
<i>Lactobacillus plantarum</i>	Productivity Ratio ≥ 0.5	Pass
<i>Lactococcus garvieae</i>	Productivity Ratio ≥ 0.5	Pass
<i>Pediococcus acidilactici</i>	Productivity Ratio ≥ 0.5	Pass
<i>Lactobacillus plantarum</i>	Productivity Ratio ≥ 0.5	Pass
<i>Citrobacter braekii</i>	No Growth	Pass
<i>Saccharomyces cerevisiae</i> ATCC 7754	No Growth	Pass
<i>Saccharomyces cerevisiae</i> ATCC 9763	No Growth	Pass
<i>Escherichia coli</i> ATCC 25922	No Growth	Pass

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Company
 3M Center, Building 275-5W-05
 St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
 Printed in USA © 3M 2022. All rights reserved.

Version 3

LAB
 Lactic Acid Bacteria Count Plate