



Certificate of Analysis

Product Name: UXL100 CLEAN-TRACE SURF ATP 100EA/PK

NEOGEN SAP Number: 700002207

SAP/ERP Number: 7100067519

3M Stock Number: GH620522827

Kit Batch Number: ...3000004595...

Date of Manufacture: ...15 Nov 2023...

Date of Expiry: ...15 Sep 2024...

Test	Temp	Inst Serial Number	Specification	Test Result
Device Background	20°C	101	Mean Bk \leq 25 RLU	Pass
Device Activity	20°C	101	Activity Mean 7109 – 7972 RLU	Pass

This is to certify that the product described above has been manufactured in accordance with 3M specifications.

AUTHORISED: (Quality Control)

DATE: ...16 Nov 2023...

Made in UK
Neogen Food Safety UK Ltd.
The Science Park
Bridgend, CF31 3NA
Tel +44 (0) 1656 641400

3M is a trademark of 3M. Used under licence in Canada
© 2021, 3M All rights reserved

QC205b UXL100 / 008 / 22.09.23 – (Copy of CONTROLLED DOCUMENT 16/11/23)