



# Petrifilm™

Created by Authorized Personnel: Jacob Billings, 2024-05-17

Manufacture Date: 2024-05-10

Expiration Date: 2025-11-08

## Product Manufacturing Certificate

## Certificate of Analysis

Product: 3M™ Petrifilm™ *Salmonella* Express Plate 6536 or 6537  
 Batch: 33XLEH  
 Stock Number: 70-2007-6998-5 or 70-2007-6999-3  
 ERP Number: 7100039574 or 7100039575

Organism Tested	Minimum Growth	Result
<i>Salmonella enterica</i> subsp <i>enterica</i> Typhimurium ATCC 14028	Typical Growth	Pass
<i>Salmonella enterica</i> subsp <i>enterica</i> Enteritidis ATCC 49223	Typical Growth	Pass
<i>Citrobacter freundii</i> FSD 548	Atypical Growth	Pass
<i>Escherichia coli</i> FSD 723	Atypical Growth	Pass

### Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Company  
 3M Center, Building 275-5W-01  
 St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.  
 Printed in USA © 3M 2022. All rights reserved.

Version 3

**SALX**  
*Salmonella* Express Plate