

# **Product Manufacturing Certificate**

## Certificate of Analysis

3M Product: 3M<sup>™</sup> Soy Protein Rapid Kit - 25 Tests per Kit, 1 Kit per Case

3M Catalog Number: L25SOY

3M Kit Lot Number: SBY1201B

3M Buffer Lot Number: SBY1201B

3M Stock Number: 70-2011-7556-2

## Performance Testing\*

### **Negative Controls**

Sample	Specification	Result
Matrix Testing	Negative, ≥6	Pass

#### **Positive Controls**

Sample	Specification	Result
Matrix Testing	Positive, ≥6	Pass

<sup>\*</sup>A minimum of four matrices are used for performance testing. All results are interpreted by eye, holding the strip an arm's length away. If the control, hook and test lines are visible the test is considered positive. If only the control and hook lines are visible the test is considered negative. If either the control or hook lines aren't visible the test is considered invalid. Positive controls are spiked at 2ppm.

This letter certifies that the product described above has been tested and found to comply with 3M's product performance specifications applicable to subject device at time of manufacture. Visit the 3M Food Safety website, www.3m.com/foodsafety for additional information. 3M is certified to ISO 9001 through an independent agency.

3M Health Care 3M Center, Building 275-5W-05 St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M is a trademark of 3M. Please recycle. Printed in USA © 3M 2020. All rights reserved.