

CERTIFICATE OF ANALYSIS



Veratox Egg Allergen
Product No.: 8450

Lot No. 3000001990
Date of Manufacture: OCT 3, 2024
Expiration Date: OCT 3, 2024

DESCRIPTION	TESTING PARAMETERS
Incubation Times	10 min, 10 min, 10 min
Standards Used	0, 2.5, 5, 10, 25 ppm
Room Temperature	21 – 23°C
Plate Reader Wavelength	650 nm


Results

TEST ITEM	SPECIFICATION	DECISION
Overall Coefficient of Variation	$\leq 10\%$	Pass
Overall Sample Deviation (Mean Overall)	$\leq 30\%$	Pass
Regression Curve (R)	≥ 0.980	Pass
Additive	≤ 1.0 ppm	Pass
Packaging	All Components Present Labeled Correctly	Pass Pass

The data above was generated using all components within the product lot.

Variances in data may occur. These are the minimum criteria when testing in a controlled setting as listed above. External factors such as stress from shipping, storage conditions, handling or contamination of components may cause loss of color activity. Neogen Corporation certifies that this test kit lot has met all internal quality control specifications for this product.

CAUTION: DO NOT USE QC RESULTS FOR DETERMINATIONS.


QC Immunodiagnostic Supervisor

Date: Wednesday, November 15, 2023



CofA-8450
Rev: 5

620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600