

CERTIFICATE OF ANALYSIS



Veratox Gliadin Allergen

SKU: 700002581

Product No.: 8480

Lot No. 3000015625

Date of Manufacture: May 21, 2024

Expiration Date: May 15, 2025

| DESCRIPTION | TESTING PARAMETERS |
|-------------------------|------------------------|
| Incubation Times | 10 min, 10 min, 10 min |
| Standards Used | 0, 5, 10, 20, 50 ppm |
| Room Temperature | 21 – 23°C |
| Plate Reader Wavelength | 650 nm |


Results

| TEST ITEM | SPECIFICATION | DECISION |
|--|------------------------|----------|
| Overall Coefficient of Variation | $\leq 10\%$ | Pass |
| Overall Sample Deviation (Mean Overall) | $\leq 30\%$ | Pass |
| Regression Curve (R) | ≥ 0.980 | Pass |
| Additive | ≤ 1.0 ppm | Pass |
| Packaging | All Components Present | Pass |
| | Labeled Correctly | Pass |

The data above was generated using all components within the product lot.

Variances in data may occur. These are the minimum criteria when testing in a controlled setting as listed above. External factors such as stress from shipping, storage conditions, handling or contamination of components may cause loss of color activity. Neogen Corporation certifies that this test kit lot has met all internal quality control specifications for this product.

CAUTION: DO NOT USE QC RESULTS FOR DETERMINATIONS.


QC Immunodiagnostic Supervisor

Date: Wednesday, June 26, 2024



CofA-700002581

Rev: 5

620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600