



Sample Collection

Product Manufacturing Certificate

Certificate of Analysis and Sterility

3M Product: 3M™ Quick Swab
3M Catalog Number: 6433
3M Lot Number: 10124
3M Stock Number: 70-2006-8489-5

Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	7.1 ± 0.3	7.45	Yes
Gamma Irradiation	25.0 kGy - 35.0 kGy	26.9 – 32.3 kGy	Yes

Terminal Irradiation Processing Parameters

This product has been subjected to Gamma irradiation as a terminal process. The minimum and maximum delivered met established specifications for this product.

This letter certifies that the product described above has been tested and found to comply with 3M's product performance specifications applicable to subject device at time of manufacture. Visit the 3M Food Safety website, www.3m.com/foodsafety for additional information. 3M is certified to ISO 9001 through an independent agency.