

CERTIFICATE OF ANALYSIS



Reveal 3-D Egg Test

SKU: 700002643

Product No.: 902082Q

Kit Lot No.: 3000018404

Device Lot No.: 3000014414

Expiration Date: Nov 02, 2025

Performance

Component	Lot Number	Expiry
Reveal 3-D Egg device	3000014414	Nov 02, 2025
Type 8 Extraction Buffer	106797	Jan 08, 2026
Swab Wetting Solution	3000019393	Jan 22, 2026

Parameter	Specification	Result
Negative (Buffer)	Test line absent. Overload and Control line present.	Pass
Positive (5 ppm spiked buffer)	Test, Overload and Control line present.	Pass
High Positive (100,000 ppm Egg)	Test line greater than or equal to Overload line. Control line present. <i>Note: Absence of Overload line acceptable.</i>	Pass

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may influence results. Neogen Corporation certifies that this lot has met all internal quality control specifications for this product.


Immunodiagnosics QC Supervisor

Date: Tuesday, August 06, 2024

CofA-700002643
Rev: 6

