

CERTIFICATE OF ANALYSIS



Milk (Casein)
Protein Rapid Kit
Item No. 700002293
Catalog No. L25MLK

Kit Lot No. MYE0103
Buffer Lot No. MYE0103
Expiration Date: 2026-03-28

Performance Testing*
Negative Controls

Sample	Specification	Result
Matrix Testing	Negative, ≥ 24	Pass

Positive Controls

Sample	Specification	Result
Matrix Testing	Positive, ≥ 24	Pass

*A minimum of four matrices are used for performance testing. All results are interpreted by eye, holding the strip an arm's length away. If the control, hook and test lines are visible the test is considered positive. If only the control and hook lines are visible the test is considered negative. If either the control or hook lines aren't visible the test is considered invalid. Positive controls are spiked at 2ppm.

This letter certifies that the product described above has been tested and found to comply with Neogen's product performance specifications applicable to subject device at time of manufacture. Visit the Neogen Food Safety website, www.Neogen.com/foodsafety for additional information. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by VanPelt Biosciences: 18211-D Flower Hill Way, Gaithersburg, MD 20879

Hannah Broucek

Hannah Broucek
Date: 2024-06-06



CofA-700002293
FS-COA-0273
Rev: 3

620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600