

CERTIFICATE OF ANALYSIS



Petrifilm® Rapid Aerobic Count Plates

Item No. 700005496 or 700005495

Catalog No. 6478 or 6479

Lot No. 418324130D

Manufacture Date:

09.05.2024

Expiration Date:

07.11.2025

ISO 1133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Productivity Ratio \geq 0.7	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	Productivity Ratio \geq 0.7	Pass
<i>Staphylococcus aureus</i> ATCC 6538 (WDCM 00032)	Productivity Ratio \geq 0.7	Pass
<i>Bacillus subtilis</i> subsp. <i>spizizenii</i> ATCC 6633 (WDCM 00003)	Productivity Ratio \geq 0.7	Pass

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured: 3M Center, Building 275-5W-05 St. Paul, MN 55144-1000

Agnieszka Górnik

Date: Wednesday, May 29, 2024



CofA-700005496

or 700005495

Rev: 1

620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200

The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600