

# CERTIFICATE OF ANALYSIS



## Reveal for Aflatoxin – Ethanol

SKU: 700002474

Product No.: 8016/8016B

Lot No. 3000013688

Device Lot No. 105270

Exp. Date: Jun 30, 2025

### Performance

TEST ITEM	SPECIFICATION	RESULT	DECISION
Sensitivity:	Detects aflatoxin at a level of 20 ppb in corn	A visible line is observed in the control zone but is absent in the test zone on the lateral flow device.	Pass
Specificity:	Does not detect non-aflatoxin substances in corn	A visible line is observed in the test and control zones on the lateral flow device.	Pass

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may influence results. Neogen Corporation certifies that this lot has met all internal quality control specifications for this product.

  
Immunodiagnosics QC Supervisor

Date: Thursday, May 02, 2024

CofA-700002474  
Rev: 3

