

CERTIFICATE OF ANALYSIS



Reveal 3-D Total Milk Test

SKU: 700002580

Product No.: 8479

Kit Lot No.: 3000019662

Device Lot No.: 3000015474

Expiration Date: Jun 13, 2025

Performance

Component	Lot Number	Expiry
Reveal 3-D Total Milk device	3000015474	Jun 13, 2025
Extraction buffer 8	106793 106795	Nov 14, 2025 Dec 12, 2025
Sterile Rayon Tipped Applicator (Swab)	3000017386	NA

Parameter	Result	Specification
Lower detectable level*	5 ppm	Test and Overload line present. Control line always present.
Upper detectable level*	>1000 ppm	Overload line less than or equal to test line. Control line always present.
Extraction buffer	Undetected	Test line absent. Overload line Present. Control line always present.
QC control	Detected	Test and Overload line present. Control line always present.

*Spiked material into extraction buffer

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may influence results. Neogen Corporation certifies that this lot has met all internal quality control specifications for this product.


Immunodiagnostics QC Supervisor

Date: Monday, July 29, 2024

