

CERTIFICATE OF ANALYSIS



Reveal Q+ for Aflatoxin

SKU: 700002497/700002498

Product No.: 8085/8085B

Kit Lot No. 3000012839

Device Lot. No. 104758

Exp. Date: AUG 31, 2025

PERFORMANCE

Incubation Time: 6 minutes

Room Temperature: ~ 30°C

AccuScan Gold



AccuScan Pro



Raptor



READERS:

Reveal AccuScan Gold

Reveal AccuScan Pro

Raptor

CURVE SETTINGS

AccuScan Gold Values:

AccuScan Pro Value:

A: -07000 B: +0887 0712109251047587.000-0.430+1.379+3.061+

C: +1639 D: +2276

TEST ITEM	SPECIFICATION	DECISION
Sensitivity:	Detects aflatoxin levels at 5, 10, 20, and 100ppb in corn.	Pass
Specificity:	Does not detect non-aflatoxin substances in corn	Pass

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may cause loss of color activity. Neogen Corporation certifies that this test kit has met all internal quality control specifications for this product.


Quality Control Supervisor

Date: Wednesday, April 24, 2024

CofA-8085

Rev: 7



CofA- 700002497

Rev: 8

620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600