CERTIFICATE OF ANALYSIS



Reveal Q+ for Aflatoxin MAX

SKU: 700002499 Product No.: 8088/8088B

Kit Lot No.	3000016676
Device Lot	No. 106444
Exp. Date:	Nov 30, 2025

PERFORMANCE Incubation Time: 6 minutes Room Temperature: ~ 30°C READERS: Reveal AccuScan Gold Reveal AccuScan Pro Raptor	AccuScan Gold	AccuScan Pro	Raptor 8088 Raptor 900 106444 Rev. 0	
CURVE SETTINGS AccuScan Gold Values: B: +0727AccuScan I 0713012251A: -07000B: +07270713012251C: +1635D: +2140	Pro Value: .064447.000-0.343+1.379+2	.943+		
TEST ITEM SPEC	IFICATION		DECISION	
~	ts aflatoxin levels at 20, 100 and 300ppb n		Pass	
~ F · · · · · · · · · · · · · · · · · ·	not detect non-aflatoxin nces in corn		Pass	

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may cause loss of color activity. Neogen Corporation certifies that this test kit has met all internal quality control specifications for this product.

Immunodiagnostics QC Supervisor

Date: Thursday, June 13, 2024

