

CERTIFICATE OF ANALYSIS



**Petrifilm® Environmental
Listeria Plates**
Item No. 700002233 or 700002232
Catalog No. 6447 or 6448

Lot No. 418324145A
Manufacture Date:
24.05.2024
Expiration Date:
22.11.2025

ISO 1133:2014 Performance Testing

| Organism Tested | Growth Specification | Result |
|-------------------------------------------------|----------------------------|--------|
| <i>Listeria monocytogenes</i> SLR-22 | Typical Listeria Growth | Pass |
| <i>Listeria innocua</i> ATCC 33090 | Typical Listeria Growth | Pass |
| <i>Enterococcus faecalis</i> ATCC 14506 | No Typical Listeria Growth | Pass |
| <i>Bacillus atrophaeus (subtilis)</i> ATCC 9372 | No Typical Listeria Growth | Pass |

Contamination Check

| Test Plan | Result |
|-----------------------------------------------------------------------|--------|
| 96 Randomized plates per batch tested per a statistical sampling plan | Pass |

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured: 3M Center, Building 275-5W-05 St. Paul, MN 55144-1000

Jarosław Matczak

Date: Thursday, June 13, 2024



CofA-700002233
or 700002232
Rev: 1

620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600