

CERTIFICATE OF ANALYSIS



Clean-Trace® Surface ATP Swab Test

Item No. 700002207
Catalog No. UXL100

Kit Lot No: 3000020434
Date of Manufacture: 05 Aug 2024
Expiration Date: 05 Jun 2025

Test	Temp	Inst Serial Number	Specification	Test Result
Device Background	20°C	293	Mean Bk \leq 25 RLU	Pass
Device Activity	20°C	293	Activity Mean 7109 – 7972 RLU	Pass

Comments:

This is to certify that the product described above has been manufactured in accordance with Neogen specifications.

Barbara Lyczko
Date: 06 Aug 2024

Manufactured in the UK by: Neogen Food Safety UK Ltd
The Science Park
Bridgend CF31 3NA
Tel +44 (0) 1656 641400



620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600