

## Clean-Trace® Water Plus - Total ATP Test


Item No. 700002100  
Catalog No. AQT200

Kit Lot No: 3000020264  
Date of Manufacture: 24 Jul 2024  
Expiration Date: 24 Jul 2025

Test	Temperature	Inst Serial Number	Specification	Test Result
Device Background	20°C	293	Mean Bk $\leq$ 20 RLU	Pass
Device Activity	20°C	293	Activity Mean 15000 – 20000 RLU	Pass

Comments:

This is to certify that the product described above has been manufactured in accordance with Neogen specifications.



Barbara Lyczko  
Date: 31 Jul 2024

Manufactured in the UK by: Neogen Food Safety UK Ltd  
The Science Park  
Bridgend CF31 3NA  
Tel +44 (0) 1656 641400



620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200  
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600