

# CERTIFICATE OF ANALYSIS



## Reveal 3-D Egg Test

SKU: 700002643

Product No.: 902082Q

Kit Lot No.: 3000020903

Device Lot No.: 3000013167

Expiration Date: Nov 02, 2025

### Performance

Component	Lot Number	Expiry
Reveal 3-D Egg device	3000013167	Nov. 02, 2025
Type 8 Extraction Buffer	106796 106797	Dec. 01, 2025 Jan. 08, 2026
Swab Wetting Solution	3000019393	Jan. 22, 2026

  

Parameter	Specification	Result
Negative (Buffer)	Test line absent. Overload and Control line present.	Pass
Positive (5 ppm spiked buffer)	Test, Overload and Control line present.	Pass
High Positive (100,000 ppm Egg)	Test line greater than or equal to Overload line. Control line present. <i>Note: Absence of Overload line acceptable.</i>	Pass

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may influence results. Neogen Corporation certifies that this lot has met all internal quality control specifications for this product.

  
Immunodiagnostics QC Supervisor

Date: Friday, August 16, 2024

CofA-700002643  
Rev: 6

