

CERTIFICATE OF ANALYSIS



Reveal 3-D Total Milk Test

SKU: 700002580

Product No.: 8479

Kit Lot No.: 3000043026

Device Lot No.: 3000041371

Expiration Date: 2026-11-21


Performance

Component	Lot Number	Expiry
Reveal 3-D Total Milk device	3000041371	2026-11-21
Extraction buffer 8	122505 119627	2027-04-06 2027-03-15
Sterile Rayon Tipped Applicator (Swab)	3000041830	2030-10-01

Parameter	Result	Specification
Lower detectable level*	5 ppm	Test and Overload line present. Control line always present.
Upper detectable level*	>1000 ppm	Overload line less than or equal to test line. Control line always present.
Extraction buffer	Undetected	Test line absent. Overload line Present. Control line always present.
QC control	Detected	Test and Overload line present. Control line always present.

*Spiked material into extraction buffer

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may influence results. Neogen Corporation certifies that this lot has met all internal quality control specifications for this product.


Immunodiagnosics QC Supervisor

Date: Monday, December 15, 2025

