

# CERTIFICATE OF ANALYSIS



## Reveal 3-D Egg Test

SKU: 700002643

Product No.: 902082Q

Kit Lot No.: 3000044649

Device Lot No.: 3000042798

Expiration Date: 2027-04-06

### Performance

Component	Lot Number	Expiry
Reveal 3-D Egg device	3000042798	2027-04-06
Type 8 Extraction Buffer	122505	2027-04-06
Swab Wetting Solution	3000042233	2027-05-26

  

Parameter	Specification	Result
Negative (Buffer)	Test line absent. Overload and Control line present.	Pass
Positive (5 ppm spiked buffer)	Test, Overload and Control line present.	Pass
High Positive (100,000 ppm Egg)	Test line greater than or equal to Overload line. Control line present. <i>Note: Absence of Overload line acceptable.</i>	Pass

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may influence results. Neogen Corporation certifies that this lot has met all internal quality control specifications for this product.

  
Quality Control Supervisor

Date: Wednesday, January 21, 2026

CofA-700002643  
Rev: 6



620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200  
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600