

# CERTIFICATE OF ANALYSIS



**Petrifilm® Rapid Yeast  
and Mold Count Plates**  
Item No. 700002138 or 700002148  
Catalog No. 6475 or 6477

**Lot No.** 3483AR  
**Manufacture Date:** 2026-01-23  
**Expiration Date:** 2027-07-24

Organism Tested	Growth Specification	Result
<i>Geotrichum candidum</i> ATCC 34614	Productivity Ratio $\geq 0.5$	Pass
<i>Candida albicans</i> ATCC 10231	Productivity Ratio $\geq 0.5$	Pass
<i>Enterococcus faecalis</i> ATCC 14506	No Growth	Pass
<i>Escherichia coli</i> ATCC 25922	No Growth	Pass

## Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	6.5 $\pm$ 1.0	6.38	Yes

## Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Solventum, 601 22nd Ave. South, Brookings, SD 57006

*Gabrielle Jackson*

Gabrielle Jackson  
Date: 2026-02-10

